



**Idaho State Police Forensic Services**

**CERTIFICATE OF ANALYSIS**

The Idaho State Police Forensic Services hereby certifies that Alcohol Simulator Solution Lot Number 09802 to be used to conduct calibration checks within the State of Idaho in accordance with the policies and/or procedures promulgated by the Department governing breath alcohol examinations. This lot has a target value of .083 with a range of 0.075 to 0.091 grams of ethyl alcohol/210 liters of vapor.

The expiration date for this lot number is March 30, 2011 at 11:59 PM.

April 27, 2009  
Date

  
Forensic Scientist

STATE OF IDAHO    )  
                              ) ss.  
County of Ada )

On this 27<sup>th</sup> of April, in the year 2009, before me, Judy Packer, a notary public, personally appeared Darren B. Jewkes, known to me to be the person whose name is subscribed to the within instrument as a Forensic Scientist for the Idaho State Police Forensic Services, and acknowledge to me that he executed the same as such Scientist.

 Notary Public

My Commission Expires: 4 October 2013

